

计量作业指引流程-电脑版

当计量人员接到派工的工单后，可以通过计量作业模块记录具体操作。

计量作业流程

1.内部人员接单

2.预约计量时间

3.签到

4.取得认证



一、进入维保作业中计量作业界面。

The screenshot displays the 'Medical Equipment Management Platform' (医疗设备管理平台) interface. The left sidebar contains navigation options: '首页' (Home), '维保作业' (Maintenance Work), '维修作业' (Repair Work), '保养作业' (Maintenance Work), '计量作业' (Measurement Work), and '不良事件管理' (Adverse Event Management). Red arrows and circles highlight the '维保作业' and '计量作业' options. The main content area shows the breadcrumb path: '当前位置 > 维保作业 > 计量作业 > 计量单管理' (Current Location > Maintenance Work > Measurement Work > Measurement Order Management). Below this, there are tabs for '待接单' (Orders to be received), '待预约' (Orders to be booked), '待签到' (Orders to be signed in), '待完工' (Orders to be completed), '待评价' (Orders to be evaluated), and '完结归档' (Completed and Archived). A search bar with the text '请输入关键词' (Please enter keywords) and a '查询' (Search) button is present. A table lists measurement orders with columns for '序号' (Serial Number), '操作' (Action), '工单号' (Order Number), '报单时间' (Order Time), '设备名称' (Equipment Name), '设备类型' (Equipment Type), '计量类型' (Measurement Type), '紧急程度' (Urgency), and '工单状态' (Order Status). The first row shows an order with ID 'JL20220117150917289' for a '裂隙灯' (Slit Lamp) on '2022/01/17 15:09:17', categorized as '内部计量' (Internal Measurement) with a '一般' (General) urgency, and a status of '待接单' (Orders to be received). The table includes '接单' (Accept Order) and '详情' (Details) buttons for the first row. At the bottom right, there are pagination controls: '上一页' (Previous Page), '1' (Page 1), '下一页' (Next Page), and '10条/页' (10 items per page). A back arrow is visible in the bottom right corner of the main content area.

医疗设备管理平台

医疗机构测试1

消息 李四

当前位置 > 维保作业 > 计量作业 > 计量单管理

计量单管理

待接单 待预约 待签到 待完工 待评价 完结归档

请输入关键词 查询

序号	操作	工单号	报单时间	设备名称	设备类型	计量类型	紧急程度	工单状态
1	接单 详情	JL20220117150917289	2022/01/17 15:09:17	裂隙灯		内部计量	一般	待接单

上一页 1 下一页 10条/页

二、点击待接单页面的接单键。

内部工程师只能看到派单给自己的工单。该工单会自动带出已录入的设备信息、报修人员所上传的附件等资料。

The screenshot displays the '医疗设备管理平台' (Medical Equipment Management Platform) interface. The left sidebar contains navigation options: '首页' (Home), '维保作业' (Maintenance Work), '维修作业' (Repair Work), '保养作业' (Maintenance Work), '计量作业' (Measurement Work), and '不良事件管理' (Adverse Event Management). The main content area is titled '医疗机构测试1' (Medical Institution Test 1) and shows the '计量单管理' (Measurement Order Management) section. The current status is '待接单' (Waiting for Order), indicated by a red '1' above the button. Below the status bar is a search input field with the placeholder '请输入关键词' (Please enter keywords) and a '查询' (Search) button. A table lists measurement orders with columns for '序号' (Serial Number), '操作' (Action), '工单号' (Work Order Number), '计量人' (Measurement Person), '报单时间' (Reporting Time), '设备名称' (Equipment Name), '设备类型' (Equipment Type), '计量类型' (Measurement Type), '紧急程度' (Urgency), and '工单状态' (Work Order Status). The first row shows a work order for '裂隙灯' (Slit Lamp) with status '待接单'. The '操作' column for this row contains a '接单' (Accept Order) button, which is highlighted with a red box and a red arrow pointing to it, labeled with a red '3'. The table also includes pagination controls at the bottom right, showing '上一页' (Previous Page), '1', '下一页' (Next Page), and '10条/页' (10 items/page).

序号	操作	工单号	计量人	报单时间	设备名称	设备类型	计量类型	紧急程度	工单状态
1	接单	JL20220117150917289	李四	2022/01/17 15:09:17	裂隙灯		内部计量	一般	待接单

三、跳转至计量单流程页面，点击下方确认接单键。

医疗设备管理平台

医疗机构测试1

消息 李四

当前位置 > 维保作业 > 计量作业 > 计量单流程

计量单号: JL20220117150917289 设备名称: 裂隙灯 唯一码: 000019 权属: 眼科

刷新

设备计量信息

报单时间: 2022/01/17 15:09:17 报单人: 张三 计量类型: 内部计量

紧急程度: 一般 图片信息: [Image] 附件信息: [Image]

下载 下载

计量内容: 一般紧急的裂隙灯内部计量

计量人信息

计量人: 李四 计量科室: 设备科

4

确认接单

四、点击页面右上角**预约**键，选择计量方式和开始计量的时间，然后点击页面下方的**确认预约**。

医疗设备管理平台

医疗机构测试1

消息 李四

当前位置 > 维保作业 > 计量作业 > 计量单详情

计量单进度 计量单详情

计量单号: JL20220117150917289 设备名称: 裂隙灯 唯一码: 000019 权属: 眼科

取消接单 预约 打印

基础信息

设备名称:	裂隙灯	设备唯一码:	000019	设备类型:	
设备所属科室:	眼科				

设备计量信息

报单时间:	2022/01/17 15:09:17	报单人:	张三	计量类型:	内部计量
紧急程度:	一般	图片信息:	 下载	附件信息:	 下载
计量内容:	一般紧急的裂隙灯内部计量				

计量人信息

计量人:	李四	计量科室:	设备科
------	----	-------	-----

首页

维保作业

维修作业

保养作业

计量作业

不良事件管理

当前位置 > 维保作业 > 计量作业 > 计量单流程

计量单号: JL20220117150917289 设备名称: 裂隙灯 唯一码: 000019 权属: 眼科

刷新

设备计量信息

报单时间: 2022/01/17 15:09:17

报单人: 张三

计量类型: 内部计量

紧急程度: 一般

图片信息:



下载

附件信息:



下载

计量内容: 一般紧急的裂隙灯内部计量

计量人信息

计量人: 李四

计量科室: 设备科

计量方式信息

计量方式 内部计量

预约开始时间 2022-01-18 16:12:41

6

确认预约



五、工程师到达现场后，点击右上角**签到**键，填写**实际开工时间**后，点击页面下方的**确认签到**。

The screenshot displays the 'Medical Equipment Management Platform' (医疗设备管理平台) interface. The left sidebar contains navigation options: 首页 (Home), 维保作业 (Maintenance Work), 维修作业 (Repair Work), 保养作业 (Maintenance Work), 计量作业 (Measurement Work), and 不良事件管理 (Adverse Event Management). The main content area is titled '医疗机构测试1' (Medical Institution Test 1) and shows the breadcrumb path: 当前位置 > 维保作业 > 计量作业 > 计量单详情 (Current Location > Maintenance Work > Measurement Work > Measurement Single Details). The page features two tabs: '计量单进度' (Measurement Single Progress) and '计量单详情' (Measurement Single Details). The measurement single number is JL20220117150917289, the device name is 裂隙灯 (Slit Lamp), the unique code is 000019, and the department is 眼科 (Ophthalmology). The '签到' (Check-in) button is highlighted with a red box and a circled '7' above it, with a red arrow pointing to it. Other buttons include '取消预约' (Cancel Appointment) and '打印' (Print). The '基础信息' (Basic Information) section shows: 设备名称: 裂隙灯, 设备唯一码: 000019, 设备类型: (blank), 设备所属科室: 眼科. The '设备计量信息' (Device Measurement Information) section shows: 报单时间: 2022/01/17 15:09:17, 报单人: 张三, 计量类型: 内部计量, 紧急程度: 一般, 图片信息: (with a download icon), 附件信息: (with a download icon). The '计量内容' (Measurement Content) is 一般紧急的裂隙灯内部计量. The '计量人信息' (Measurement Person Information) section shows: 计量人: 李四, 计量科室: 设备科.

首页

维保作业

维修作业

保养作业

计量作业

不良事件管理

当前位置 > 维保作业 > 计量作业 > 计量单流程

计量单号: JL20220117150917289 设备名称: 裂隙灯 唯一码: 000019 权属: 眼科

刷新

报单时间: 2022/01/17 15:09:17

报单人: 张三

计量类型: 内部计量

紧急程度: 一般

图片信息:



下载

附件信息:



下载

计量内容: 一般紧急的裂隙灯内部计量

计量人信息

计量人: 李四

计量科室: 设备科

计量方式信息

计量方式: 内部计量

预约开始时间: 2022/01/18 16:12:41

签到信息

* 实际开工时间

2022-01-18 16:15:25



8

确认签到



六、完成设备计量工作后，点击右上角**完工**键。填写**实际完工时间**、**完工工时**、**完工情况**等信息后，点击页面下方的**确认取得认证**。设备计量后的图片、视频等文件，可以作为凭证上传为**附件**。此时该计量工单完成，待管理者评价。

The screenshot displays the 'Medical Equipment Management Platform' interface. On the left is a dark blue sidebar with navigation options: '首页' (Home), '维保作业' (Maintenance Work), '维修作业' (Repair Work), '保养作业' (Maintenance Work), '计量作业' (Measurement Work), and '不良事件管理' (Adverse Event Management). The '计量作业' option is currently selected.

The main content area is titled '医疗机构测试1' (Medical Institution Test 1) and shows the breadcrumb path: '当前位置 > 维保作业 > 计量作业 > 计量单详情' (Current Location > Maintenance Work > Measurement Work > Measurement Order Details). Below this, there are tabs for '计量单进度' (Measurement Order Progress) and '计量单详情' (Measurement Order Details), with the latter being active.

The measurement order details include:

- 计量单号: JL20220117150917289
- 设备名称: 裂隙灯
- 唯一码: 000019
- 权属: 眼科

At the top right of the details section, there are three buttons: '取消签到' (Cancel Sign-in), '完工' (Completion), and '打印' (Print). The '完工' button is highlighted with a red box and a red arrow pointing to it from a red circle containing the number '9'.

The details are organized into sections:

- 基础信息** (Basic Information):
 - 设备名称: 裂隙灯
 - 设备唯一码: 000019
 - 设备类型:
 - 设备所属科室: 眼科
- 设备计量信息** (Equipment Measurement Information):
 - 报单时间: 2022/01/17 15:09:17
 - 报单人: 张三
 - 计量类型: 内部计量
 - 紧急程度: 一般
 - 图片信息: [Image icon] 下载
 - 附件信息: [Image icon] 下载
 - 计量内容: 一般紧急的裂隙灯内部计量
- 计量人信息** (Measurement Personnel Information):
 - 计量人: 李四
 - 计量科室: 设备科

当前位置 > 维保作业 > 计量作业 > 计量单流程

计量单号: JL20220117150917289 设备名称:裂隙灯 唯一码: 000019 权属: 眼科

刷新

签到信息

* 实际开工时间: 2022/01/18 16:15:25

计量情况

* 实际完工时间 2022-01-18 16:18:32 * 计量工时 (小时) 0.05

认证附件:  + 上传
下载 删除

认证图片:  + 上传
下载 删除

* 计量完成情况 顺利

计量备注

10

确认取得认证

